



## CSIR-Indian Institute of Petroleum

(Council of Scientific & Industrial Research)  
P.O.I.I.P., MOHKAMPUR, DEHRADUN – 248005 (UA) INDIA  
Ph.0135-2525762 & 2525754, Fax. 0135-2660072, 2660202-203

E-Mail- [cosp@iip.res.in](mailto:cosp@iip.res.in)  
Website : [www.iip.res.in](http://www.iip.res.in)



### CORRIGENDUM

**Tender Notice Ref.: IIP/PUR/1/19-20/481/GN/ASD(1) ICPOES**

**SUB: Amendment in Specification for Inductively Coupled Plasma Optical Emission Spectrometer (ICP-OES)**

With reference to above tender for Inductively Coupled Plasma Optical Emission Spectrometer (ICP-OES), please note that **we have amend the specification** are as under:-

Page No 54, Point No 8, Sub point :12 : Nitrogen Generator with warranty

<u>EXISTING</u>	<u>TO BE READ AS</u>
<u>CHAPTER 4</u>  Nitrogen Generator with warranty	<u>CHAPTER 4</u>  <b>Nitrogen generator with warranty:</b>  Standard nitrogen generator should be quoted with following or better specifications: Flow: 750 ml/min, Purity: 99.999%, Delivery pressure up to 5 barg (75 psig), Non condensing and oil free air.

All other entries, terms & conditions are same.

  
28/1  
(Controller of Stores & Purchase)