



**CSIR – INDIAN INSTITUTE OF PETROLEUM
(COUNCIL OF SCIENTIFIC & INDUSTRIAL RESEARCH)
HARIDWAR ROAD, MOHKAMPUR, DEHRADUN INDIA – 248005**

**EXPRESSION OF INTEREST
for
FINDING / SEARCHING DOMESTIC MANUFACTURERS OR SUPPLIERS FOR
SCIENTIFIC / R&D ITEMS**

IIP REF: IIP/EOI/2020/MII/

HARD COPY of your proposal to be submitted to:

**Controller of Stores & Purchase
CSIR - INDIAN INSTITUTE OF PETROLEUM
(COUNCIL OF SCIENTIFIC & INDUSTRIAL RESEARCH)
HARIDWAR ROAD, MOHKAMPUR, DEHRADUN INDIA – 248005
E-Mail: cosp@iip.res.in, rawats@iip.res.in
Website: <http://www.iip.res.in>**

LAST DATE FOR SUBMITTING WRITTEN PROPOSAL: Dec 1 , 2020, 3.00 P.M

**ENQUIRIES SEEKING NAMES OF INDNETORS / END USERS SHALL NOT BE ENTERTAINED. ALL
INSTRUCTIONS ARE AVAILABLE IN THIS DOCUMENT ONLY. PLEASE READ IT THOROUGHLY.**

CSIR-Indian Institute of Petroleum (IIP), Dehradun, an ISO 9001 Institute, is one of the leading constituent laboratories under Council of Scientific & Industrial Research (CSIR) engaged in R&D work in petroleum refining, natural gas and petro-chemicals and contributing towards creation of state of the art technology & products intends to procure various scientific instruments/ items as per the detailed specifications provided in this document. Normally most of these items with the kind of specifications and parameters required for users in IIP for their research purposes are available through international manufacturers / suppliers. In pursuance of recent policies / guideline /orders issued by various Ministries of the Govt. of India to encourage domestic manufacturing and preference to domestic suppliers under various programs like 'Make in India', 'Atma Nirbhar Bharat' etc. interested firms capable of offering the listed items are requested to respond to this notice.

Firms are requested to refer to the **Order Nos. P-45021/2/2017-PP (BE-II) dt. 15.06.2017 as amended vide order of even number 28.05.2018, 29.05.2019, 04.06.2020, and 16.09.2020** issued by Public Procurement Section of DPIIT, Min. of Commerce & Industry, Government of India in their own interest to know about the provisions related to domestic suppliers for participation in open tenders. Firms may also refer to various other policies / programs of the Govt. related to promoting domestic manufacturing and/or supply.

Following is the list of items with their detailed specifications that are required by users in CSIR-IIP to meet their research requirements. WE STRONGLY ENCOURAGE INDIAN MANUFACTURERS / SUPPLIERS OF SAME / SIMILAR ITMES TO GO THROUGH THE TECHNICAL PARAMETERS AND WRITE TO US EXPRESSING THEIR INTEREST TO MANUFACTURE / SUPPLY THE SAME. WHEREVER IT IS FELT THAT ANY / SOME PARAMETERS ARE RESTRICTIVE OR SUPERFLOUS IN NATURE, DOMESTIC MANUFACTURERS CAN BRING THE SAME TO OUR NOTICE.

DOMESTIC MANUFACTURERS / SUPPLIERS, INDIAN SUBSIDIARIES OR AGENTS OF FOREIGN MANUFACTURERS / SUPPLIERS ARE REQUESTED TO GO THROUG THE PROVISION OF ORDERS ISSUED BY DPIIT, MIN. OF COMMERCE & INDUSTRY OR ANY OTHER ORDER / CIRCULAR ISSUED BY OTHER MINISITRIES TO EXPRESS THEIR INTEREST TO SUPPLY THE ITEMS REQUIRED BY CSIR-IIP IN COMPLIANCE OF THE PROVISIONS OF THE THOSE ORDER(S)/CIRCULARS.

AFTER EVALUATING RESPONSE TO THIS EOI NOTICE, CSIR-IIP WILL DECIDE ABOUT FLOATING GLOBAL TENDER ENQUIRY (GTE) AFTER SEEKING APPROVAL OF THE COMPETENT AUTHORITY ONLY FOR SUCH ITEMS FOR WHICH NO DOMESTIC MANUFACTURER OR SUPPLIER HAS EXPRESSED INTEREST TO SUPPLY THE ITMES WITH THE REQUIRED SPECIFICATIONS IN PURSUANCE OF THE SAID GOVT. CIRCULARS / ORDERS / NOTIFICATIONS.

CRITICAL DATE SHEET

क्रमांक Sl. No.	प्रक्रमStage	दिनांक और समय Date & Time
1.	प्रकाशनकादिनांक और समय Publish Date & Time	06.11.2020 1600hrs IST
2.	दस्तावेजविक्रय / डाउनलोडकाप्रारंभदिनांक और समय Sale /document Download Start Date & Time	06.11.2020 1600hrs IST
3.	संदेह / सवालपूछनेकीअंतिमदिनांक और समय Last Date & time for receipt of queries	13.11.2020 1500hrs IST
4.	बोलीपूर्वसम्मेलन, यदिहोतो Pre-bid Conference, if any	NOT APPLICABLE
5.	बोलीजमाकरनेकीप्रारंभिकदिनांक और समय Bid Submission Start Date & time	06.11.2020 1600hrs IST
6.	बोलीजमाकरनेकीअंतिमदिनांक और समय Bid Submission End Date & Time	01.12.2020 1500hrs IST
7.	बोलीखोलेजानेकीदिनांक और समय Bid Opening Date & Time	02.12.2020 1500 HRS IST

EXPRESSION OF INTEREST

1	Item Name :	Automatic ASTM D86 Distillation unit:
	<p>Tentative Specifications</p> <p>The automated ASTM D86 distillation unit would be required for analyzing the distillation characteristics of the liquid fractions obtained from waste plastics conversion. Auto Distillation Units should be a compact self-contained Instruments with liquid coolant free cooling system, delivered with all accessories to immediately begin testing.</p> <p>Brief Specification:</p> <p>Technical Specification for Auto Distillation Equipment As per test method ASTM D 86 D850, D1078, IP 123 and ISO 3405. It should include: distillation flask (50-100 mL), receiver, vapor probe with centering device, heater plates, Flask connection silicon tubes, receiver cap and condenser cleaner.</p> <p>User Interface: Digital user interface with alpha Numeric key pad with dedicated function keys and solvent proof protection. The distillation analyzer should have connectivity ports such as USB ports, HDMI and Bluetooth/Wifi to connect to mouse, keyboard, printer and monitor without any need of a PC.</p> <p>Flask Heating System: Low mass and low voltage non-contact heater with self-positioning heating system and automatic heater lifting and lowering. Heating coil should not be in contact with the flask to avoid any damage to the flask. Fully automatic initial heat settings, distillation rate regulation and final heating regulation, to start test select the group and push start button.</p> <p>Condenser Temperature range 0 to 80°C programmable constant temperature, temperature ramping or special temperature profile, instantaneously ready at switch on, automatic cooling via pielter cooling and switch between groups within 5 mins.</p> <p>Receiver chamber should be capable of maintaining programmable temperature from 0 to 60°C and also should measure the charge volume before distillation.</p> <p>The analyzer should have facility to detect temperature of the flask and LED system to indicate the operator.</p> <p>Instrument should be able to perform manual distillation with ASTM thermometer and manual heating control.</p> <p>The offered instrument should have a built-in calibrated atmospheric pressure sensor and humidity sensor.</p> <p>The offered analyzer will have Safety interlock for the receiver chamber; door movement should be automatic without any manual intervention.</p>	

For and on behalf of CSIR

Sd/-

Controller of Stores & Purchase

IMPORTANT INSTRUCTIONS

- INTERESTED FIRMS NEED TO SUBMIT PROPOSALS IN HARD COPY FORMAT LATEST **BY Dec.1, 2020**. WE MAY NOT BE ABLE TO CONSIDER PROPOSALS RECEIVED AFTER THIS DATE. THIS WILL NOT HOWEVER PREVENT YOU FROM PARTICIPATING IN THE OPEN TENDERING PROCESS, IF DONE SUBSEQUENTLY THROUGH GeM OR CPPP PORTAL.
- PLEASE NOTE THAT CSIR-IIP'S ITEM SPECIFICATIONS / PARAMETERS ARE BASED ON USERS RESEARCH REQUIREMENTS AND FUNCTIONAL NEEDS AND MAY INVITE THE FIRMS FOR PRE INDENT MEETING / PRESENTATION /DISCUSSIONS. INTIMATION RE. THAT SHALL BE SENT SEPARATELY BY EMAIL ONLY AND ALSO HOSTED ON CSIR - IIP WEBSITE. THERE IS NO NEED TO FOLLOW UP TELEPHONICALLY AND/OR THROUGH EMAIL.
- WRITE ON THE MAIN ENVELOPE 'EXPRESSION **OF INTEREST**' FOR 'ITEM NAME' WITH 'REF. NO.' AS GIVEN IN THIS BID DOCUMENT. THE ENVELOPE MUST BE ADDRESSED TO THE **CONTROLLER OF STORES & PURCHASE** AND SENT TO THE ADDRESS PROVIDED ON THE TITLE PAGE OF THIS DOCUMENT.
- SEND ONLY TECHNICAL LITERATURE / BROCHURE. THERE IS NO NEED TO ENCLOSE RECOMMENDATION LETTERS OR ANY OTHER DETAILS NOT CONCERNED WITH THE TECHNICAL ASPECT OF THE ITEM AT THIS STAGE.**PLEASE NOTE THAT NO FORMAL OPENING WILL TAKE PLACE AT A SPECIFIED DATE OR TIME FOR THE OFFERS RECEIVED AS THIS BEING PROSPECTING ACTIVITY IS JUST AN INFORMATION GATHERING / SEEKING EXERCISE AND NOT A TENDERING PROCESS.**
- THIS EXERCISE IS BEING CARRIED OUT TO GET DETAILED IDEA ABOUT THE TYPE OF SPECIFICATIONS, TECHNICAL PARAMETERS INDIAN MANUFACTURERS ARE WILLING AND ABLE TO MANUFACTURE AND / OR SUPPLY (IN PURSUANCE OF VARIOUS GOVERNMENT POLICIES / GUIDELINES / INSTRUCTIONS) FOR THE TYPE OF EQUIPMENT DESIRED TO BE PROCURED BY CSIR – IIP SO THAT CSIR – IIP CAN AVOID GOING FOR GLOBAL TENDERING AND ENCOURAGE DOMESTIC MANUFACTURING AND COMPETITION.
- **BEING A PUBLIC FUNDED ORGANIZATION CSIR-IIP IS ENTITLED TO CONCESSIONAL GST @5% AND CONCESSIONAL CUSTOMS DUTY PAYABLE UNDER...**
- OPTIONALLY FIRMS MAY ALSO ENCLOSE 'BUDGETARY OFFER' TO ENABLE US TO GET BROAD IDEA ABOUT THE PRICING. **PLEASE NOTE THAT THIS IS NOT A TENDER OR BIDDING PROCESS AND THE PRICES SENT BY YOU SHALL NOT BE USED TO MAKE PURCHASE DECISION AT ALL.** WE WILL FRAME ITEM SPECIFICATIONS & DETAILED BID DOCUMENT AFTER WHICH TENDERING PROCESS SHALL BEGIN. YOU MAY GET OPPORTUNITY TO PARTICIPATE IN OUR TENDERING PROCESS WHERE BID PRICES ARE ALSO REQUIRED TO BE QUOTED. **DISCLOSING BUDGETARY PRICES AT THIS STAGE IS PURELY OPTIONAL & FIRMS MAY SKIP THIS IF THEY FEAR LOSING COMPETITIVE ADVANTAGE AT THE TENDERING STAGE LATER.**
- WARRANTY, EXTENDED WARRANTY AND AMC OFFERINGS AVAILABLE (PERIOD OF COVERAGE AND ESTIMATE) SHOULD BE PROVIDED FOR ALL ITEMS.

For and on behalf of CSIR
Sd/-

CONTROLLER OF STORES & PURCHASE

