



सी०एस०आइ०आर०-भारतीय पेट्रोलियम संस्थान

(वैज्ञानिक एवं औद्योगिक अनुसंधान परिषद)

CSIR-Indian Institute of Petroleum

(Council of Scientific & Industrial Research)

Institute

P.O.I.I.P., MOHKAMPUR, DEHRADUN – 248005 (UA) INDIA

Ph.0135-2525762 & 2525754, Fax. 0135-2660072, 2660202-203

E-Mail- cosp@iip.res.in, rawats@iip.res.in

Website : www.iip.res.in



Creating
Future
Fuels

ISO-9001-2015 Regd.

10.03.2021

CORRIGENDUM

Tender Notice Ref.: IIP/PUR/1/2021/RRB/488/13180PO:

Tender ID: 2021_CSIR_71459_1

Sub: Correction in specification

With reference to above tender for High Pressure Fixed Bed Micro Reactor please note that following corrections in the Tender specifications as detailed below:

1) Reactor Dimension:

- 1.0 Reactor dimension : OD 20mm X ID 12 mm X L 660mm or better thickness
- 2.0 Reactor material INC 600 with thermocouple fitted
- 3.0 Heating zone : 350 mm

2) Furnace : spilt furnace single zone

Max operating temperature : 1000 OC

operating voltage : 220 VAC

3) MFC : Inlet pressure : 80 bar out let pressure : 70

Flow rate : 0-200SCCM

All other terms & conditions remains same.

For & on behalf of CSIR

(Controller of Stores & Purchase)